



## 830-2013 ADDENDUM 2

### SUPPLY AND DELIVERY OF AUTOMATED EXTERNAL DEFIBRILLATORS

#### **URGENT**

**PLEASE FORWARD THIS DOCUMENT TO  
WHOEVER IS IN POSSESSION OF THE BID  
OPPORTUNITY**

ISSUED: March 24, 2014  
BY: B Dupuis Smith  
TELEPHONE NO. (204) 986-2492

**THIS ADDENDUM SHALL BE INCORPORATED  
INTO THE BID OPPORTUNITY AND SHALL  
FORM A PART OF THE CONTRACT  
DOCUMENTS**

Template Version: A20131129

---

**Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 8 of Form A: Bid may render your Bid non-responsive.**

---

#### **PART D – SUPPLEMENTAL CONDITIONS**

Revise D13.1 to read: Notwithstanding C11, the warranty period shall begin on the date of Total Performance and shall expire five (5) years thereafter unless extended pursuant to C11.3, in which case it shall expire when provided for thereunder.

#### **PART E – SPECIFICATIONS**

Add: E2.2(g) to read: The unit noted in E2.2 shall include all items required for use and intended purpose. Items shall include but are not limited to: case, batteries or battery packs (provided by manufacturer), pads (both adult and pediatric), electrodes etc.