



977-2014 ADDENDUM 1

SUPPLY AND DELIVERY OF POWER COTS

URGENT

**PLEASE FORWARD THIS DOCUMENT TO
WHOEVER IS IN POSSESSION OF THE BID
OPPORTUNITY**

ISSUED: October 16, 2014
BY: B Dupuis Smith
TELEPHONE NO. (204) 986-2492

**THIS ADDENDUM SHALL BE INCORPORATED
INTO THE BID OPPORTUNITY AND SHALL
FORM A PART OF THE CONTRACT
DOCUMENTS**

Template Version: A20131129

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 9 of Form A: Bid may render your Bid non-responsive.

PART E – SPECIFICATIONS

Revise: E2.2(a)(i) to read: Battery power supply changeover will not necessitate any type of electrical power cabling being disconnected and reconnected, Battery unit must be able to be **changed** when cot is at any height with patient at any position.