



756-2018 ADDENDUM 1

PROVISION OF LABORATORY TESTING FOR CITY POOLS

URGENT

**PLEASE FORWARD THIS DOCUMENT TO
WHOEVER IS IN POSSESSION OF THE BID
OPPORTUNITY**

ISSUED: October 15, 2018
BY: Glen Stefanysheh
TELEPHONE NO. 204 986-2638

**THIS ADDENDUM SHALL BE INCORPORATED
INTO THE BID OPPORTUNITY AND SHALL
FORM A PART OF THE CONTRACT
DOCUMENTS**

Template Version: A20160708

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 8 of Form A: Bid may render your Bid non-responsive.

PART E – SPECIFICATIONS

Revise: E2.7 (d) to read: provide test results turnaround time within **three** Calendar Days.

Delete: E2.1 (a) Shipping costs (if any);

QUESTIONS AND ANSWERS

Q1: In Specifications E2.7(d) it specifies to provide results within 2 calendar days. For the analysis of Pseudomonas what method of analysis are you requiring? Some methods do not allow for reporting of confirmatory results in that time frame (can report presumptive results but not confirmatory).

A1: **Refer to 756-2018 Addendum 1 E2.7 (d), revised to three Calendar Days. The method of analysis required for pseudomonas is APHA 9213E, this is a non-proprietary industry standard test.**

Q2: In Specifications E2.1(a) Shipping costs (if any) – to what shipping costs is this referring? The specifications indicates that city staff will deliver samples and pick up supplies so it is not clear what shipping costs this is referring to. Please clarify.

A2: **There are no shipping costs.**