



770-2019 ADDENDUM 1

REQUEST FOR PROPOSAL FOR SUPPLY AND INSTALLATION OF GAS CHROMATOGRAPH

URGENT

PLEASE FORWARD THIS DOCUMENT TO WHOEVER IS IN POSSESSION OF THE TENDER

ISSUED: June 1, 2020
BY: Brock Bradford
TELEPHONE NO. 204 986-4752

THIS ADDENDUM SHALL BE INCORPORATED INTO THE TENDER AND SHALL FORM A PART OF THE CONTRACT DOCUMENTS

Template Version: A20160708

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Tender, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 10 of Form A: Bid may render your Bid non-responsive.

PART D – SUPPLEMENTAL CONDITIONS

Revise D9.4: to read:

D9.4 Training shall be completed fifteen (15) **Business Days** post installation as per requirements **in** E2.9.

Delete: D15.1

Revise: D16.1 to read:

D16.1 Notwithstanding C11, the warranty period shall begin on the date of Total Performance and shall expire after one year thereafter unless extended pursuant to C11.3, in which case it shall expire when provided for thereunder.

PART E – SPECIFICATIONS

Revise: E2.3 (e) to read:

E2.3(e) operates under the latest available version of applicable software.

Revise: E2.4 (j) to read:

E2.4(j) have GC dimensions not exceeding 2 inches of the following dimension, 27" x 20" x 20" (W x H x D). **This does not include the Injector Tower Automatic Liquid Sampler;**

Revise E2.6 (e): to read:

E2.6(e) include at least all the following parameters with a MDL of zero point zero five (0.05) ug/mL (micrograms per milliliter) for each parameter for water, wastewater and industrial waste matrices ;

- (i) Acetic Acid
- (ii) Propionic Acid
- (iii) i-Butyric Acid
- (iv) n-Butyric Acid
- (v) i-Valeric Acid
- (vi) n-Valeric Acid

- (vii) i-Caproic Acid
- (viii) n-Caproic Acid
- (ix) Heptanoic Acid