



770-2019 ADDENDUM 2

REQUEST FOR PROPOSAL FOR SUPPLY AND INSTALLATION OF GAS CHROMATOGRAPH

URGENT

PLEASE FORWARD THIS DOCUMENT TO WHOEVER IS IN POSSESSION OF THE TENDER

ISSUED: June 2, 2020
BY: Brock Bradford
TELEPHONE NO. 204 986-4752

THIS ADDENDUM SHALL BE INCORPORATED INTO THE TENDER AND SHALL FORM A PART OF THE CONTRACT DOCUMENTS

Template Version: A20190115

Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Tender, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 10 of Form A: Bid may render your Bid non-responsive.

PART D – SUPPLEMENTAL CONDITIONS

Revise: D2.2(b) to read:

D2.2(b) Supply and installation of the required software.

PART E – SPECIFICATIONS

Revise: E2.4(h) to read:

E2.4(h) have wide temperature operating range from **ambient temperature** to four hundred and fifty(450) degree Celsius and with a minimum ramp temperature of one (1) degree Celsius;

Revise: E2.9(c) to read:

E2.9(c) provide on-site training to cover the GC-FID Software Data analysis, GC-FID reporting and GC-FID VFA analysis for a minimum of three (3) staff. **If additional workstations are required for training**, the rental and set-up of computers with the GC-FID software is to be provided by the vendor;